



May 08, 2024

Stephen Herda
North Dakota Army National Guard - DoD
Attn: Environmental Office
Bldg 30 Fraine Barracks Dr.
Bismarck, ND 58504

RE: Project: DoD_PFAS 537.1
Pace Project No.: 35876724

Dear Stephen Herda:

Enclosed are the analytical results for sample(s) received by the laboratory on May 02, 2024. The results relate only to the samples included in this report. Results reported herein conform to the DOD Quality Systems Manual Version 5.4 and the laboratory's Quality Assurance Manual, where applicable, unless otherwise noted in the body of the report.

The test results provided in this final report were generated by each of the following laboratories within the Pace Network:

- Pace Analytical Services - Ormond Beach

If you have any questions concerning this report, please feel free to contact me.

Sincerely,

A handwritten signature in blue ink, appearing to read "Lorraine".

Lorraine Noronha
lorraine.noronha@pacelabs.com
(386)672-5668
Project Manager

Enclosures



REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



CERTIFICATIONS

Project: DoD_PFAS 537.1

Pace Project No.: 35876724

Pace Analytical Services Ormond Beach

8 East Tower Circle, Ormond Beach, FL 32174

Alaska DEC- CS/UST/LUST

Alabama Certification #: 41320

California Certification# 3096

Colorado Certification: FL NELAC Reciprocity

Connecticut Certification #: PH-0216

Delaware Certification: FL NELAC Reciprocity

DoD-ANAB #:ADE-3199

Florida Certification #: E83079

Georgia Certification #: 955

Guam Certification: FL NELAC Reciprocity

Hawaii Certification: FL NELAC Reciprocity

Illinois Certification #: 200068

Indiana Certification: FL NELAC Reciprocity

Kansas Certification #: E-10383

Kentucky Certification #: 90050

Louisiana Certification #: FL NELAC Reciprocity

Louisiana Environmental Certificate #: 05007

Maine Certification #: FL01264

Maryland Certification: #346

Massachusetts Certification #: M-FL1264

Michigan Certification #: 9911

Mississippi Certification: FL NELAC Reciprocity

Missouri Certification #: 236

Montana Certification #: Cert 0074

Nebraska Certification: NE-OS-28-14

New Hampshire Certification #: 2958

New Jersey Certification #: FL022

New York Certification #: 11608

North Carolina Environmental Certificate #: 667

North Carolina Certification #: 12710

North Dakota Certification #: R-216

Ohio DEP 87780

Oklahoma Certification #: D9947

Pennsylvania Certification #: 68-00547

Puerto Rico Certification #: FL01264

South Carolina Certification: #96042001

Tennessee Certification #: TN02974

Texas Certification: FL NELAC Reciprocity

US Virgin Islands Certification: FL NELAC Reciprocity

Virginia Environmental Certification #: 460165

West Virginia Certification #: 9962C

Wisconsin Certification #: 399079670

Wyoming (EPA Region 8): FL NELAC Reciprocity

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



SAMPLE SUMMARY

Project: DoD_PFAS 537.1

Pace Project No.: 35876724

Lab ID	Sample ID	Matrix	Date Collected	Date Received
35876724001	DIX 240430 DW01	Drinking Water	04/30/24 09:15	05/02/24 10:40
35876724002	FRB 1	Drinking Water	04/30/24 09:12	05/02/24 10:40
35876724003	WSH 240430 DW01	Drinking Water	04/30/24 13:25	05/02/24 10:40
35876724004	FRB 2	Drinking Water	04/30/24 13:22	05/02/24 10:40
35876724005	LIS 240430 DW01	Drinking Water	04/30/24 15:20	05/02/24 10:40
35876724006	FRB 3	Drinking Water	04/30/24 15:18	05/02/24 10:40
35876724007	WAH 240430 DW01	Drinking Water	04/30/24 16:53	05/02/24 10:40
35876724008	FRB 4	Drinking Water	04/30/24 16:50	05/02/24 10:40
35876724009	DLR 240501 DW01	Drinking Water	05/01/24 11:00	05/02/24 10:40
35876724010	FRB 5	Drinking Water	05/01/24 10:57	05/02/24 10:40

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



SAMPLE ANALYTE COUNT

Project: DoD_PFAS 537.1

Pace Project No.: 35876724

Lab ID	Sample ID	Method	Analysts	Analytes Reported	Laboratory
35876724001	DIX 240430 DW01	EPA 537.1	TMM1	22	PASI-O
35876724002	FRB 1	EPA 537.1	TMM1	22	PASI-O
35876724003	WSH 240430 DW01	EPA 537.1	TMM1	22	PASI-O
35876724004	FRB 2	EPA 537.1	TMM1	22	PASI-O
35876724005	LIS 240430 DW01	EPA 537.1	TMM1	22	PASI-O
35876724006	FRB 3	EPA 537.1	TMM1	22	PASI-O
35876724007	WAH 240430 DW01	EPA 537.1	TMM1	22	PASI-O
35876724008	FRB 4	EPA 537.1	TMM1	22	PASI-O
35876724009	DLR 240501 DW01	EPA 537.1	TMM1	22	PASI-O
35876724010	FRB 5	EPA 537.1	TMM1	22	PASI-O

PASI-O = Pace Analytical Services - Ormond Beach

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



PROJECT NARRATIVE

Project: DoD_PFAS 537.1

Pace Project No.: 35876724

Method: EPA 537.1

Description: 537.1 PFAS Compounds, DOD

Client: NDARNG - North Dakota Army National Guard - DoD

Date: May 08, 2024

General Information:

10 samples were analyzed for EPA 537.1 by Pace Analytical Services Ormond Beach. All samples were received in acceptable condition with any exceptions noted below or on the chain-of custody and/or the sample condition upon receipt form (SCUR) attached at the end of this report.

Hold Time:

The samples were analyzed within the method required hold times with any exceptions noted below.

Sample Preparation:

The samples were prepared in accordance with EPA 537.1 with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):

All criteria were within method requirements with any exceptions noted below.

Continuing Calibration:

All criteria were within method requirements with any exceptions noted below.

Internal Standards:

All internal standards were within QC limits with any exceptions noted below.

Surrogates:

All surrogates were within QC limits with any exceptions noted below.

Method Blank:

All analytes were below the report limit in the method blank, where applicable, with any exceptions noted below.

Laboratory Control Spike:

All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:

All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

Additional Comments:

This data package has been reviewed for quality and completeness and is approved for release.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: DoD_PFAS 537.1

Pace Project No.: 35876724

Sample: DIX 240430 DW01 Lab ID: 35876724001 Collected: 04/30/24 09:15 Received: 05/02/24 10:40 Matrix: Drinking Water

Parameters	Results	Units	DL	LOD	LOQ	DF	Prepared	Analyzed	CAS No.	Qual
537.1 PFAS Compounds, DOD										
Analytical Method: EPA 537.1 Preparation Method: EPA 537.1										
Initial Volume/Weight: 262.79 mL Final Volume/Weight: 1 mL										
Pace Analytical Services - Ormond Beach										
11CI-PF3OUdS	0.0018U	ug/L	0.00090	0.0018	0.0038	1	05/03/24 04:18	05/03/24 23:03	763051-92-9	
9CI-PF3ONS	0.0018U	ug/L	0.00054	0.0018	0.0038	1	05/03/24 04:18	05/03/24 23:03	756426-58-1	
ADONA	0.0018U	ug/L	0.00070	0.0018	0.0038	1	05/03/24 04:18	05/03/24 23:03	919005-14-4	
HFPO-DA	0.0019U	ug/L	0.00071	0.0019	0.0038	1	05/03/24 04:18	05/03/24 23:03	13252-13-6	
NEtFOSAA	0.0019U	ug/L	0.00091	0.0019	0.0038	1	05/03/24 04:18	05/03/24 23:03	2991-50-6	
NMeFOSAA	0.0019U	ug/L	0.00078	0.0019	0.0038	1	05/03/24 04:18	05/03/24 23:03	2355-31-9	
PFBS	0.0017U	ug/L	0.00080	0.0017	0.0038	1	05/03/24 04:18	05/03/24 23:03	375-73-5	
PFDA	0.0019U	ug/L	0.00074	0.0019	0.0038	1	05/03/24 04:18	05/03/24 23:03	335-76-2	
PFHxA	0.0019U	ug/L	0.00067	0.0019	0.0038	1	05/03/24 04:18	05/03/24 23:03	307-24-4	
PFDoA	0.0019U	ug/L	0.00083	0.0019	0.0038	1	05/03/24 04:18	05/03/24 23:03	307-55-1	
PFHpA	0.0019U	ug/L	0.00095	0.0019	0.0038	1	05/03/24 04:18	05/03/24 23:03	375-85-9	
PFHxS	0.0017U	ug/L	0.00085	0.0017	0.0038	1	05/03/24 04:18	05/03/24 23:03	355-46-4	
PFNA	0.0019U	ug/L	0.00060	0.0019	0.0038	1	05/03/24 04:18	05/03/24 23:03	375-95-1	
PFOS	0.0018U	ug/L	0.00071	0.0018	0.0038	1	05/03/24 04:18	05/03/24 23:03	1763-23-1	
PFOA	0.0019U	ug/L	0.00058	0.0019	0.0038	1	05/03/24 04:18	05/03/24 23:03	335-67-1	
PFTeDA	0.0019U	ug/L	0.00085	0.0019	0.0038	1	05/03/24 04:18	05/03/24 23:03	376-06-7	
PFTrDA	0.0019U	ug/L	0.00085	0.0019	0.0038	1	05/03/24 04:18	05/03/24 23:03	72629-94-8	
PFUnA	0.0019U	ug/L	0.00071	0.0019	0.0038	1	05/03/24 04:18	05/03/24 23:03	2058-94-8	
Surrogates										
13C2-PFDA (S)	92	%		70-130		1	05/03/24 04:18	05/03/24 23:03		
13C2-PFHxA (S)	91	%		70-130		1	05/03/24 04:18	05/03/24 23:03		
NEtFOSAA-d5 (S)	84	%		70-130		1	05/03/24 04:18	05/03/24 23:03		
HFPO-DAS (S)	81	%		70-130		1	05/03/24 04:18	05/03/24 23:03		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: DoD_PFAS 537.1

Pace Project No.: 35876724

Sample: FRB 1 Lab ID: 35876724002 Collected: 04/30/24 09:12 Received: 05/02/24 10:40 Matrix: Drinking Water

Parameters	Results	Units	DL	LOD	LOQ	DF	Prepared	Analyzed	CAS No.	Qual
537.1 PFAS Compounds, DOD										
Analytical Method: EPA 537.1 Preparation Method: EPA 537.1										
Initial Volume/Weight: 256.98 mL Final Volume/Weight: 1 mL										
Pace Analytical Services - Ormond Beach										
11CI-PF3OUdS	0.0018U	ug/L	0.00092	0.0018	0.0039	1	05/03/24 04:18	05/04/24 03:47	763051-92-9	
9CI-PF3ONS	0.0018U	ug/L	0.00055	0.0018	0.0039	1	05/03/24 04:18	05/04/24 03:47	756426-58-1	
ADONA	0.0018U	ug/L	0.00072	0.0018	0.0039	1	05/03/24 04:18	05/04/24 03:47	919005-14-4	
HFPO-DA	0.0019U	ug/L	0.00073	0.0019	0.0039	1	05/03/24 04:18	05/04/24 03:47	13252-13-6	
NEtFOSAA	0.0019U	ug/L	0.00093	0.0019	0.0039	1	05/03/24 04:18	05/04/24 03:47	2991-50-6	
NMeFOSAA	0.0019U	ug/L	0.00080	0.0019	0.0039	1	05/03/24 04:18	05/04/24 03:47	2355-31-9	
PFBS	0.0018U	ug/L	0.00082	0.0018	0.0039	1	05/03/24 04:18	05/04/24 03:47	375-73-5	
PFDA	0.0019U	ug/L	0.00076	0.0019	0.0039	1	05/03/24 04:18	05/04/24 03:47	335-76-2	
PFHxA	0.0019U	ug/L	0.00068	0.0019	0.0039	1	05/03/24 04:18	05/04/24 03:47	307-24-4	
PFDoA	0.0019U	ug/L	0.00085	0.0019	0.0039	1	05/03/24 04:18	05/04/24 03:47	307-55-1	
PFHpA	0.0019U	ug/L	0.00097	0.0019	0.0039	1	05/03/24 04:18	05/04/24 03:47	375-85-9	
PFHxS	0.0018U	ug/L	0.00087	0.0018	0.0039	1	05/03/24 04:18	05/04/24 03:47	355-46-4	
PFNA	0.0019U	ug/L	0.00061	0.0019	0.0039	1	05/03/24 04:18	05/04/24 03:47	375-95-1	
PFOS	0.0018U	ug/L	0.00073	0.0018	0.0039	1	05/03/24 04:18	05/04/24 03:47	1763-23-1	
PFOA	0.0019U	ug/L	0.00059	0.0019	0.0039	1	05/03/24 04:18	05/04/24 03:47	335-67-1	
PFTeDA	0.0019U	ug/L	0.00087	0.0019	0.0039	1	05/03/24 04:18	05/04/24 03:47	376-06-7	
PFTrDA	0.0019U	ug/L	0.00087	0.0019	0.0039	1	05/03/24 04:18	05/04/24 03:47	72629-94-8	
PFUnA	0.0019U	ug/L	0.00073	0.0019	0.0039	1	05/03/24 04:18	05/04/24 03:47	2058-94-8	
Surrogates										
13C2-PFDA (S)	96	%		70-130		1	05/03/24 04:18	05/04/24 03:47		
13C2-PFHxA (S)	93	%		70-130		1	05/03/24 04:18	05/04/24 03:47		
NEtFOSAA-d5 (S)	82	%		70-130		1	05/03/24 04:18	05/04/24 03:47		
HFPO-DAS (S)	87	%		70-130		1	05/03/24 04:18	05/04/24 03:47		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: DoD_PFAS 537.1

Pace Project No.: 35876724

QC Batch:	1008833	Analysis Method:	EPA 537.1
QC Batch Method:	EPA 537.1	Analysis Description:	537.1 PFOA Compounds, DOD
		Laboratory:	Pace Analytical Services - Ormond Beach

Associated Lab Samples: 35876724001, 35876724002, 35876724003, 35876724004, 35876724005, 35876724006

METHOD BLANK: 5544620 Matrix: Water
 Associated Lab Samples: 35876724001, 35876724002, 35876724003, 35876724004, 35876724005, 35876724006

Parameter	Units	Blank Result	DL	LOD	LOQ	Analyzed	Qualifiers
11CI-PF3OUdS	ug/L	0.0019U	0.00095	0.0019	0.0040	05/03/24 22:16	
9CI-PF3ONS	ug/L	0.0019U	0.00057	0.0019	0.0040	05/03/24 22:16	
ADONA	ug/L	0.0019U	0.00074	0.0019	0.0040	05/03/24 22:16	
HFPO-DA	ug/L	0.0020U	0.00075	0.0020	0.0040	05/03/24 22:16	
NEtFOSAA	ug/L	0.0020U	0.00096	0.0020	0.0040	05/03/24 22:16	
NMeFOSAA	ug/L	0.0020U	0.00082	0.0020	0.0040	05/03/24 22:16	
PFBS	ug/L	0.0018U	0.00084	0.0018	0.0040	05/03/24 22:16	
PFDA	ug/L	0.0020U	0.00078	0.0020	0.0040	05/03/24 22:16	
PFDaA	ug/L	0.0020U	0.00087	0.0020	0.0040	05/03/24 22:16	
PFHpA	ug/L	0.0020U	0.0010	0.0020	0.0040	05/03/24 22:16	
PFHxA	ug/L	0.0020U	0.00070	0.0020	0.0040	05/03/24 22:16	
PFHxS	ug/L	0.0018U	0.00089	0.0018	0.0040	05/03/24 22:16	
PFNA	ug/L	0.0020U	0.00063	0.0020	0.0040	05/03/24 22:16	
PFOA	ug/L	0.0020U	0.00061	0.0020	0.0040	05/03/24 22:16	
PFOS	ug/L	0.0019U	0.00075	0.0019	0.0040	05/03/24 22:16	
PFTeDA	ug/L	0.0020U	0.00089	0.0020	0.0040	05/03/24 22:16	
PFTrDA	ug/L	0.0020U	0.00089	0.0020	0.0040	05/03/24 22:16	
PFUnA	ug/L	0.0020U	0.00075	0.0020	0.0040	05/03/24 22:16	
13C2-PFDA (S)	%	96		70-130		05/03/24 22:16	
13C2-PFHxA (S)	%	94		70-130		05/03/24 22:16	
HFPO-DAS (S)	%	88		70-130		05/03/24 22:16	
NEtFOSAA-d5 (S)	%	85		70-130		05/03/24 22:16	

LABORATORY CONTROL SAMPLE: 5544621

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
11CI-PF3OUdS	ug/L	0.15	0.13	86	70-130	
9CI-PF3ONS	ug/L	0.15	0.13	89	70-130	
ADONA	ug/L	0.15	0.14	91	70-130	
HFPO-DA	ug/L	0.16	0.14	87	70-130	
NEtFOSAA	ug/L	0.16	0.12	75	70-130	
NMeFOSAA	ug/L	0.16	0.14	85	70-130	
PFBS	ug/L	0.14	0.13	94	70-130	
PFDA	ug/L	0.16	0.15	92	70-130	
PFDaA	ug/L	0.16	0.14	87	70-130	
PFHpA	ug/L	0.16	0.15	92	70-130	
PFHxA	ug/L	0.16	0.15	92	70-130	
PFHxS	ug/L	0.15	0.14	93	70-130	
PFNA	ug/L	0.16	0.15	94	70-130	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
 without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: DoD_PFAS 537.1

Pace Project No.: 35876724

LABORATORY CONTROL SAMPLE: 5544621

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
PFOA	ug/L	0.16	0.15	94	70-130	
PFOS	ug/L	0.15	0.13	91	70-130	
PFTeDA	ug/L	0.16	0.13	83	70-130	
PFTTrDA	ug/L	0.16	0.13	82	70-130	
PFUnA	ug/L	0.16	0.14	88	70-130	
13C2-PFDA (S)	%			96	70-130	
13C2-PFHxA (S)	%			96	70-130	
HFPO-DAS (S)	%			88	70-130	
NEtFOSAA-d5 (S)	%			79	70-130	

LABORATORY CONTROL SAMPLE: 5544622

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
11CI-PF3OUdS	ug/L	0.0019	0.0016J	86	50-150	
9CI-PF3ONS	ug/L	0.0019	0.0017J	91	50-150	
ADONA	ug/L	0.0019	0.0017J	89	50-150	
HFPO-DA	ug/L	0.002	0.0017J	84	50-150	
NEtFOSAA	ug/L	0.002	0.0017J	86	50-150	
NMeFOSAA	ug/L	0.002	0.0018J	92	50-150	
PFBS	ug/L	0.0018	0.0017J	95	50-150	
PFDA	ug/L	0.002	0.0019J	94	50-150	
PFDaA	ug/L	0.002	0.0017J	85	50-150	
PFHpA	ug/L	0.002	0.0018J	90	50-150	
PFHxA	ug/L	0.002	0.0018J	92	50-150	
PFHxS	ug/L	0.0018	0.0017J	96	50-150	
PFNA	ug/L	0.002	0.0019J	95	50-150	
PFOA	ug/L	0.002	0.0020J	99	50-150	
PFOS	ug/L	0.0019	0.0017J	93	50-150	
PFTeDA	ug/L	0.002	0.0016J	80	50-150	
PFTTrDA	ug/L	0.002	0.0016J	80	50-150	
PFUnA	ug/L	0.002	0.0018J	89	50-150	
13C2-PFDA (S)	%			97	70-130	
13C2-PFHxA (S)	%			95	70-130	
HFPO-DAS (S)	%			88	70-130	
NEtFOSAA-d5 (S)	%			83	70-130	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 5544633 5544634

Parameter	Units	MS 35876724001		MSD 5544634		MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
		Result	Spike Conc.	Spike Conc.	Result								
11CI-PF3OUdS	ug/L	0.0018U	0.15	0.15	0.13	0.13	89	85	70-130	5	30		
9CI-PF3ONS	ug/L	0.0018U	0.15	0.15	0.13	0.12	88	85	70-130	4	30		
ADONA	ug/L	0.0018U	0.15	0.15	0.13	0.13	91	87	70-130	4	30		
HFPO-DA	ug/L	0.0019U	0.16	0.16	0.13	0.13	82	81	70-130	1	30		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: DoD_PFAS 537.1

Pace Project No.: 35876724

Parameter	Units	35876724001		5544633		5544634		% Rec	% Rec	% Rec	Limits	RPD	Max RPD	Qual
		Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result								
NEtFOSAA	ug/L	0.0019U	0.16	0.16	0.12	0.12	79	74	70-130	6	30			
NMeFOSAA	ug/L	0.0019U	0.16	0.16	0.14	0.13	89	85	70-130	5	30			
PFBS	ug/L	0.0017U	0.14	0.14	0.13	0.12	91	83	70-130	9	30			
PFDA	ug/L	0.0019U	0.16	0.16	0.14	0.14	92	88	70-130	4	30			
PFDoA	ug/L	0.0019U	0.16	0.16	0.14	0.13	91	84	70-130	8	30			
PFHpA	ug/L	0.0019U	0.16	0.16	0.14	0.14	92	87	70-130	6	30			
PFHxA	ug/L	0.0019U	0.16	0.16	0.14	0.13	91	85	70-130	6	30			
PFHxS	ug/L	0.0017U	0.14	0.14	0.13	0.12	92	87	70-130	6	30			
PFNA	ug/L	0.0019U	0.16	0.16	0.15	0.15	95	93	70-130	2	30			
PFOA	ug/L	0.0019U	0.16	0.16	0.15	0.14	94	91	70-130	3	30			
PFOS	ug/L	0.0018U	0.14	0.14	0.13	0.12	92	85	70-130	8	30			
PFTeDA	ug/L	0.0019U	0.16	0.16	0.14	0.13	87	85	70-130	3	30			
PFTrDA	ug/L	0.0019U	0.16	0.16	0.14	0.13	87	82	70-130	7	30			
PFUnA	ug/L	0.0019U	0.16	0.16	0.14	0.14	91	87	70-130	5	30			
13C2-PFDA (S)	%						99	92	70-130					
13C2-PFHxA (S)	%						94	89	70-130					
HFPO-DAS (S)	%						87	84	70-130					
NEtFOSAA-d5 (S)	%						85	80	70-130					

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: DoD_PFAS 537.1

Pace Project No.: 35876724

QC Batch:	1009642	Analysis Method:	EPA 537.1
QC Batch Method:	EPA 537.1	Analysis Description:	537.1 PFOA Compounds, DOD
		Laboratory:	Pace Analytical Services - Ormond Beach

Associated Lab Samples: 35876724007, 35876724008, 35876724009, 35876724010

METHOD BLANK: 5549243 Matrix: Water

Associated Lab Samples: 35876724007, 35876724008, 35876724009, 35876724010

Parameter	Units	Blank Result	DL	LOD	LOQ	Analyzed	Qualifiers
11CI-PF3OUdS	ug/L	0.0019U	0.00095	0.0019	0.0040	05/08/24 00:54	
9CI-PF3ONS	ug/L	0.0019U	0.00057	0.0019	0.0040	05/08/24 00:54	
ADONA	ug/L	0.0019U	0.00074	0.0019	0.0040	05/08/24 00:54	
HFPO-DA	ug/L	0.0020U	0.00075	0.0020	0.0040	05/08/24 00:54	
NEtFOSAA	ug/L	0.0020U	0.00096	0.0020	0.0040	05/08/24 00:54	
NMeFOSAA	ug/L	0.0020U	0.00082	0.0020	0.0040	05/08/24 00:54	
PFBS	ug/L	0.0018U	0.00084	0.0018	0.0040	05/08/24 00:54	
PFDA	ug/L	0.0020U	0.00078	0.0020	0.0040	05/08/24 00:54	
PFDaA	ug/L	0.0020U	0.00087	0.0020	0.0040	05/08/24 00:54	
PFHpA	ug/L	0.0020U	0.0010	0.0020	0.0040	05/08/24 00:54	
PFHxA	ug/L	0.0020U	0.00070	0.0020	0.0040	05/08/24 00:54	
PFHxS	ug/L	0.0018U	0.00089	0.0018	0.0040	05/08/24 00:54	
PFNA	ug/L	0.0020U	0.00063	0.0020	0.0040	05/08/24 00:54	
PFOA	ug/L	0.0020U	0.00061	0.0020	0.0040	05/08/24 00:54	
PFOS	ug/L	0.0019U	0.00075	0.0019	0.0040	05/08/24 00:54	
PFTeDA	ug/L	0.0020U	0.00089	0.0020	0.0040	05/08/24 00:54	
PFTTrDA	ug/L	0.0020U	0.00089	0.0020	0.0040	05/08/24 00:54	
PFUnA	ug/L	0.0020U	0.00075	0.0020	0.0040	05/08/24 00:54	
13C2-PFDA (S)	%	97		70-130		05/08/24 00:54	
13C2-PFHxA (S)	%	99		70-130		05/08/24 00:54	
HFPO-DAS (S)	%	90		70-130		05/08/24 00:54	
NEtFOSAA-d5 (S)	%	95		70-130		05/08/24 00:54	

LABORATORY CONTROL SAMPLE: 5549244

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
11CI-PF3OUdS	ug/L	0.0076	0.0073	97	70-130	
9CI-PF3ONS	ug/L	0.0075	0.0075	100	70-130	
ADONA	ug/L	0.0076	0.0075	99	70-130	
HFPO-DA	ug/L	0.008	0.0074	93	70-130	
NEtFOSAA	ug/L	0.008	0.0068	85	70-130	
NMeFOSAA	ug/L	0.008	0.0072	90	70-130	
PFBS	ug/L	0.0071	0.0068	96	70-130	
PFDA	ug/L	0.008	0.0078	98	70-130	
PFDaA	ug/L	0.008	0.0069	86	70-130	
PFHpA	ug/L	0.008	0.0079	99	70-130	
PFHxA	ug/L	0.008	0.0079	98	70-130	
PFHxS	ug/L	0.0073	0.0074	101	70-130	
PFNA	ug/L	0.008	0.0080	100	70-130	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: DoD_PFAS 537.1

Pace Project No.: 35876724

LABORATORY CONTROL SAMPLE: 5549244

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
PFOA	ug/L	0.008	0.0083	103	70-130	
PFOS	ug/L	0.0074	0.0075	101	70-130	
PFTeDA	ug/L	0.008	0.0068	85	70-130	
PFTTrDA	ug/L	0.008	0.0068	85	70-130	
PFUnA	ug/L	0.008	0.0074	92	70-130	
13C2-PFDA (S)	%			104	70-130	
13C2-PFHxA (S)	%			105	70-130	
HFPO-DAS (S)	%			95	70-130	
NEtFOSAA-d5 (S)	%			94	70-130	

LABORATORY CONTROL SAMPLE: 5549245

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
11Cl-PF3OUdS	ug/L	0.0019	0.0014J	72	50-150	
9Cl-PF3ONS	ug/L	0.0019	0.0014J	73	50-150	
ADONA	ug/L	0.0019	0.0015J	77	50-150	
HFPO-DA	ug/L	0.002	0.0015J	74	50-150	
NEtFOSAA	ug/L	0.002	0.0013J	67	50-150	
NMeFOSAA	ug/L	0.002	0.0014J	68	50-150	
PFBS	ug/L	0.0018	0.0013J	72	50-150	
PFDA	ug/L	0.002	0.0015J	73	50-150	
PFDaA	ug/L	0.002	0.0014J	69	50-150	
PFHpA	ug/L	0.002	0.0015J	75	50-150	
PFHxA	ug/L	0.002	0.0016J	78	50-150	
PFHxS	ug/L	0.0018	0.0015J	82	50-150	
PFNA	ug/L	0.002	0.0015J	77	50-150	
PFOA	ug/L	0.002	0.0016J	81	50-150	
PFOS	ug/L	0.0019	0.0015J	81	50-150	
PFTeDA	ug/L	0.002	0.0014J	68	50-150	
PFTTrDA	ug/L	0.002	0.0014J	69	50-150	
PFUnA	ug/L	0.002	0.0015J	73	50-150	
13C2-PFDA (S)	%			100	70-130	
13C2-PFHxA (S)	%			104	70-130	
HFPO-DAS (S)	%			95	70-130	
NEtFOSAA-d5 (S)	%			97	70-130	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 5549246 5549247

Parameter	Units	MS 35876724007		MSD		MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
		Result	Conc.	Spike Conc.	Spike Conc.								
11Cl-PF3OUdS	ug/L	0.0018U	0.0072	0.0071	0.0067	0.0066	93	92	70-130	2	30		
9Cl-PF3ONS	ug/L	0.0018U	0.0072	0.007	0.0067	0.0069	94	97	70-130	2	30		
ADONA	ug/L	0.0018U	0.0072	0.0071	0.0072	0.0071	99	99	70-130	1	30		
HFPO-DA	ug/L	0.0019U	0.0077	0.0075	0.0070	0.0067	91	89	70-130	4	30		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: DoD_PFAS 537.1

Pace Project No.: 35876724

Parameter	Units	5549246		5549247		MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	Max RPD	Qual
		35876724007 Result	MS Spike Conc.	MSD Spike Conc.	MS Result							
NEtFOSAA	ug/L	0.0019U	0.0077	0.0075	0.0060	0.0062	79	82	70-130	3	30	
NMeFOSAA	ug/L	0.0019U	0.0077	0.0075	0.0064	0.0068	83	89	70-130	6	30	
PFBS	ug/L	0.0017U	0.0068	0.0067	0.0067	0.0065	98	98	70-130	2	30	
PFDA	ug/L	0.0019U	0.0077	0.0075	0.0072	0.0072	93	95	70-130	0	30	
PFDoA	ug/L	0.0019U	0.0077	0.0075	0.0066	0.0065	86	86	70-130	1	30	
PFHpA	ug/L	0.0019U	0.0077	0.0075	0.0075	0.0075	94	94	70-130	1	30	
PFHxA	ug/L	0.0019U	0.0077	0.0075	0.0077	0.0073	100	96	70-130	5	30	
PFHxS	ug/L	0.0017U	0.007	0.0069	0.0070	0.0070	100	101	70-130	1	30	
PFNA	ug/L	0.0019U	0.0077	0.0075	0.0074	0.0076	97	101	70-130	3	30	
PFOA	ug/L	0.0019U	0.0077	0.0075	0.0079	0.0078	102	103	70-130	1	30	
PFOS	ug/L	0.0018U	0.0071	0.007	0.0069	0.0070	97	101	70-130	2	30	
PFTeDA	ug/L	0.0019U	0.0077	0.0075	0.0065	0.0066	85	87	70-130	0	30	
PFTrDA	ug/L	0.0019U	0.0077	0.0075	0.0065	0.0066	83	86	70-130	2	30	
PFUnA	ug/L	0.0019U	0.0077	0.0075	0.0067	0.0069	87	91	70-130	3	30	
13C2-PFDA (S)	%						99	101	70-130			
13C2-PFHxA (S)	%						102	103	70-130			
HFPO-DAS (S)	%						92	90	70-130			
NEtFOSAA-d5 (S)	%						90	89	70-130			

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALIFIERS

Project: DoD_PFAS 537.1

Pace Project No.: 35876724

DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.

ND - Not Detected at or above LOD.

J - The reported result is an estimated value.

LOD - Limit of Detection adjusted for dilution factor, percent moisture, initial weight and final volume.

LOQ - Limit of Quantitation adjusted for dilution factor, percent moisture, initial weight and final volume.

DL - Adjusted Method Detection Limit.

S - Surrogate

1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Analyte was not detected and is reported as less than the LOD or as defined by the customer.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.

TNI - The NELAC Institute.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

PFAS Results Reporting and Notification

What are per- and poly-fluoroalkyl substances and where do they come from?

Per- and polyfluoroalkyl substances (PFAS) are a group of thousands of man-made chemicals. PFAS have been used in a variety of industrial and consumer products around the globe, including in the U.S., for decades. Due to their widespread use and environmental persistence, most people in the U.S. have been exposed to certain PFAS. PFAS have been used to make coatings and products that are used as oil and water repellents for carpets, clothing, paper packaging for food, and cookware. They are also contained in some foams (aqueous film-forming foam or AFFF) used for fighting petroleum fires.

Is there a federal or North Dakota regulation for PFAS in drinking water?

In May 2016, the U.S. Environmental Protection Agency (EPA) established a lifetime drinking water health advisory (HA) level at 70 parts per trillion (ppt) for individual or combined concentrations of perfluorooctanoic acid (PFOA) and perfluorooctanesulfonic acid (PFOS). Both chemicals are types of PFAS. In North Dakota, there is not a PFAS drinking water regulation.

The Department of Defense (DoD) issued a policy in 2023 to monitor drinking water for PFAS at all DoD owned and operated water systems at a minimum of every two years. The DoD policy states that if water sampling results confirm that drinking water contains PFOA and PFOS at individual or combined concentrations greater than the 2016 EPA HA level of 70 ppt, water systems would 1) take immediate action to reduce exposure to PFOS or PFOA by providing alternative drinking water; and 2) evaluate and implement corrective actions to reduce levels below 70 ppt, or determine if the system should be permanently removed from use.

A final Drinking water regulation was just issued in early 2024 (<https://www.epa.gov/sdwa/and-polyfluoroalkyl-substances-pfas>). North Dakota will follow all Federal regulations as soon as they are enforceable.

What about the EPA's 2022 interim Health Advisories or proposed regulations?

EPA issued interim Health Advisories for PFOS and PFOA in 2022. However, these newer levels are below quantifiable limits (i.e., below detection levels). In March 2023, EPA announced a proposed National Primary Drinking Water Regulation (NPDWR) for six PFAS including PFOA, PFOS, PFNA, HFPO-DA (GenX Chemicals), PFHxS, and PFBS. The EPA finalized new regulations on April 10, 2024, after the public comment period and water systems will have five years to comply with the new regulations.

In anticipation of this EPA drinking water regulation and to account for emerging science that shows potential health effects of PFOS and PFOA at levels lower than 70 ppt, DoD continues to evaluate its efforts to address PFAS in drinking water, and what actions we can take to be prepared to incorporate this standard, such as reviewing our current data and collecting additional sampling where necessary. DoD remains committed to communicating and engaging with our communities throughout this process.

NDARNG efforts to ensure our Soldiers are consuming safe drinking water

At ARNG facilities where the ARNG is not the drinking water purveyor (i.e. a drinking water system that supplies water to an Army National Guard facility, but is not owned by the Army National Guard), will request the most recent finished drinking water sampling data for PFAS from the purveyor. If PFAS sampling data is not received from the purveyor within 90 days of the request, the Army National Guard facility will conduct sampling of its finished water. Upon review of this requirement, NDARNG found that five systems required sampling.

NDARNG contacted those systems and received 1) information that sampling had not been conducted or that sampling data exceeded their useful timeline and 2) permission to sample water from those systems. That sampling was accomplished on April 30 through May 1, 2024. Samples were submitted to a contract laboratory for analysis, and the results received. A copy of the results is attached to this notice.

Below MRL

The NDARNG is pleased to report that drinking water testing results were below the Minimum Reporting Level (MRL) for all 18 PFAS compounds covered by the sampling method, including PFOA and PFOS. This means that PFAS were not detected in the water system. Results are summarized below. In accordance with DoD policy, the water system will be resampled every two years for continued protection.

Analyte	Result	Status
11Cl-PF3OUdS	< 0.0018 µg/L	Does not exceed MRL
9Cl-PF3ONS	< 0.0018 µg/L	Does not exceed MRL
ADONA	< 0.0018 µg/L	Does not exceed MRL
HFPO-DA	< 0.0019 µg/L	Does not exceed MRL
NEtFOSAA	< 0.0019 µg/L	Does not exceed MRL
NMeFOSAA	< 0.0019 µg/L	Does not exceed MRL
PFBS	< 0.0017 µg/L	Does not exceed MRL
PFDA	< 0.0019 µg/L	Does not exceed MRL
PFHxA	< 0.0019 µg/L	Does not exceed MRL
PFDoA	< 0.0019 µg/L	Does not exceed MRL
PFHpA	< 0.0019 µg/L	Does not exceed MRL
PFHxS	< 0.0017 µg/L	Does not exceed MRL
PFNA	< 0.0019 µg/L	Does not exceed MRL
PFOS	< 0.0018 µg/L	Does not exceed MRL
PFOA	< 0.0019 µg/L	Does not exceed MRL
PFTeDA	< 0.0019 µg/L	Does not exceed MRL
PFTTrDA	< 0.0019 µg/L	Does not exceed MRL
PFUnA	< 0.0019 µg/L	Does not exceed MRL

[NDDEQ PFAS Fact Sheet](#)

[NDDEQ PFAS Information Webpage](#)

[EPA PFAS Information Website](#)